

Quality Control

## Certificate of Analysis

QC release date: 2021-02-24

**Product:** CK-MB FS  
**Product Code:** 11641

Cat. No.: 1 1641 02 10 026

---

**Delivery of kits:** No of kits: 14  
Kit Lot/Los: 50268435  
Kit Expiry: 2022-05-21

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent 1 Lot No.: 30261  
Expiry: 2022-05-21

Reagent 2 Lot No.: 30261  
Expiry: 2022-05-21

### Examination:

Appearance: Reagent 1: clear, colourless  
Reagent 2: clear, colourless

Odour: Reagent 1: slight smell (Thiol)  
Reagent 2: no characteristic odour

**pH:** Reagent 1 ( 25 °C ± 1 °C): 6.30  
Reagent 2 ( 25 °C ± 1 °C): 8.97  
R1+R2 ( 37 °C ± 1 °C): 6.56

**Absorbance:** Reagent 1: 0.224  
340 nm / 37 °C R1 + R2: 0.332

**Control sera:** TruLab N TruLab P  
Lot: 25894 Lot: 25895

Fresh (at 37°C) 29.7 U/L 50.8 U/L

Target values 29.8 23.8 - 35.8 50.4 40.3 - 60.4

---

### Additional remarks:

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka