

Quality Control

Certificate of Analysis

QC release date: 2020-09-23

Product: CK-MB FS
Product Code: 11641

Cat. No.: 1 1641 99 10 021

Delivery of kits: No of kits: 15
Kit Lot/Los: 60141400
Kit Expiry: 2021-12

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 29463
Expiry: 2021-12-11

Reagent 2 Lot No.: 29463
Expiry: 2021-12-11

Examination:

Appearance: Reagent 1: clear, colourless
Reagent 2: clear, colourless

Odour: Reagent 1: slight smell (Thiol)
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 6.28
Reagent 2 (25 °C ± 1 °C): 8.95
R1+R2 (37 °C ± 1 °C): 6.54

Absorbance: Reagent 1: 0.229
340 nm / 37 °C R1 + R2: 0.324

Control sera: TruLab N TruLab P
Lot: 25894 Lot: 25895

Fresh (at 37°C) 29.7 U/L 53.4 U/L

Target values 29.8 23.8 - 35.8 50.4 40.3 - 60.4

Additional remarks: Mf-Date: 11.09.2020

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka