

Quality Control

Certificate of Analysis

QC release date: 2019-12-17

Product: LDL-C Select FS
Product Code: 14121

Cat. No.: 1 4121 99 10 021

Delivery of kits: No of kits: 6
Kit Lot/Los: 60136117
Kit Expiry: 2021-05

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent 1 Lot No.: 28124
Expiry: 2021-05-31

Reagent 2 Lot No.: 28124
Expiry: 2021-05-31

Examination:

Visual check Reagent 1: clear, yellowish to slightly pink
Reagent 2: clear, yellowish

Smell Reagent 1: no characteristic odour
Reagent 2: no characteristic odour

pH: Reagent 1 (25 °C ± 1 °C): 6.78
Reagent 2 (25 °C ± 1 °C): 6.99

Absorbance: Reagent 1: 0.001
600-700 nm / 37 °C R1 + R2: 0.003

Control sera: TruLab L 1 TruLab L 2
Lot: 25174 Lot: 26040

Test results 92.7 mg/dL 131 mg/dL

Target values 95.2 76.2 - 114 137 110 - 164

Additional remarks: Mf-Date: 12.11.2019

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka