

Quality Control

## Certificate of Analysis

QC release date: 2021-04-06

**Product:** LDL-C Select FS  
**Product Code:** 14121

Cat. No.: 1 4121 99 10 021

---

**Delivery of kits:** No of kits: 12  
Kit Lot/Los: 60145834  
Kit Expiry: 2022-08-31

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent 1 Lot No.: 30070  
Expiry: 2022-08-31

Reagent 2 Lot No.: 30070  
Expiry: 2022-08-31

### Examination:

Appearance: Reagent 1: clear, yellowish to slightly pink  
Reagent 2: clear, yellowish

Odour: Reagent 1: no characteristic odour  
Reagent 2: no characteristic odour

**pH:** Reagent 1 ( 25 °C ± 1 °C): 6.77  
Reagent 2 ( 25 °C ± 1 °C): 7.00

**Absorbance:** Reagent 1: 0.001  
600-700 nm / 37 °C R1 + R2: 0.003

**Control sera:** TruLab L 1 TruLab L 2  
Lot: 27037 Lot: 27038

Test results 93.9 mg/dL 151 mg/dL

Target values 93.6 74.9 - 112 151 121 - 181

---

**Additional remarks:** Mf-Date: 02.03.2021

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka