

## CERTIFICADO DE ANÁLISE

### CERTIFICATE OF ANALYSIS

<b>MS ANVISA</b>	<b>80115310042</b>
<b>CÓDIGO / CODE</b>	<b>107095</b>
<b>COD. INTERNO / INTERNAL CODE</b>	<b>1100075K</b>
<b>DESCRIÇÃO / DESCRIPTION</b>	<b>LACTATO R1 3X20ML + R2 1X15ML</b>
	<b>KIT</b>
	LOTE / LOT: 10709520116
	DATA DE FABR./MANUF. DATE 03/2020
	DATA DE VALIDADE / EXPIRY DATE: 05/2021
	<b>REAGENTE 1 / REAGENT 1</b>
	COD. INTERNO / INTERNAL CODE: 1100020R1K
	LOTE / LOT: 1072132099
	DATA DE VALIDADE / EXPIRY DATE: 05/2021
	<b>REAGENTE 2 / REAGENT 2</b>
	COD. INTERNO / INTERNAL CODE: 1100015R2K
	LOTE / LOT: 1072142099
	DATA DE VALIDADE / EXPIRY DATE: 05/2021

DESCRIÇÃO / DESCRIPTION	ESPECIFICAÇÃO / SPECIFICATION	RESULTADO / RESULTS
APARENCIA/APPEARANCE R1	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	Conforme/According
APARENCIA/APPEARANCE R2	LIMPIDO, INCOLOR A AMARELADO - CLEAR, COLOURLESS TO YELLOWISH	Conforme/According
ODOR/ODOUR R1	ODOR CARACTERISTICO DETERGENTES - CHARACTERISTIC ODOUR DETERGENT	Conforme/According
ODOR/ODOUR R2	SEM ODOR CARACTERISTICO - NO CHARACTERISTIC ODOUR	Conforme/According
pH (25 oC) R1	8.90 - 9.10	Conforme/According
PH (25 oC) R2	2,50 - 2,70	Conforme/According
pH (25 oC) R1 + R2	8.75 - 9.15	Conforme/According
ABSORBANCIA/ABSORBANCE R1 - 340 nm	0.000 - 0.100	Conforme/According
ABSORBANCIA/ABSORBANCE R2 - 340 nm	0,000 - 0,100	Conforme/According
ABSORBANCIA/ABSORBANCE R1+R2 - 340 nm	0.000 - 0.300	Conforme/According
TRULAB N - LOTE: 25894 (mg/dL)	10.3 - 15.8	13.62
TRULAB P - LOTE: 25895 (mg/dL)	23.9 - 36.6	31.40

OBS.: Este documento foi produzido eletronicamente e é válido sem assinatura / This document has been produced eletronically and is valid without a signature.

**Monique Brasil Vianna de Sá**  
Controle de Qualidade / Quality Control