

Quality Control

Certificate of Analysis

QC release date: 2019-05-27

Product: Triglycerides FS

Product Code: 15710

Cat. No.: 1 5710 99 10 023

Delivery of kits: No of kits: 80
Kit Lot/Los: 60133193
Kit Expiry: 2021-05

Delivery of bulk: Volume:
Processing Lot:

Results for:

Reagent Lot No.: 26820
Expiry: 2021-05

Examination:

Visual check: Reagent : clear, yellow

Smell Reagent : like phenol

pH: Reagent (25 °C ± 1 °C): 7.16

Absorbance: Reagent : 0.018
546 nm/ 15-25°C

Time to reach endpoint ≤ 10 minutes at 37 °C

Control sera:	TruLab N Lot: 24382	TruLab P Lot: 24383
Fresh (at 37°C)	91.4 mg/dL	167 mg/dL
Target values	92.5 75.8 - 109	165 136 - 195

Additional remarks: Mf-Date: 16.05.2019

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka