## DiaSys Diagnostic Systems GmbH 65558 Holzheim, GERMANY

**Quality Control** 

**Product:** 

**Certificate of Analysis** 

Triglycerides FS

Product Code: 15710

Cat. No.: 1 5710 99 10 023

**Delivery of kits:** No of kits: 70

Kit Lot/Los: 60137315 Kit Expiry: 2021-12

**Delivery of bulk:** Volume:

Processing Lot:

Results for:

Reagent Lot No.: 28119

Expiry: 2021-12-08

QC release date: 2019-12-19

**Examination:** 

Visual check: Reagent: 28119

Smell Reagent: 2021-12-08

**pH:** Reagent  $(25 \text{ °C} \pm 1 \text{ °C})$ : 7.18

**Absorbance:** Reagent: 0.017

546 nm/ 15-25°C

Time to reach endpoint ≤ 10 minutes at 37 °C

Control sera: TruLab N TruLab P

Lot: 24382 Lot: 24383

Fresh (at 37°C) 92.5 mg/dL 168 mg/dL

Target values 92.5 75.8 - 109 165 136 - 195

Additional remarks: Mf-Date: 05.12.2019

This lot meets our specifications.

This certificate of analysis is a computer printout and has therefore not been signed by hand.

Sign.: Dr. G. Gorka