

Quality Control

## Certificate of Analysis

QC release date: 2019-04-08

**Product:** Lp(a) 21 FS  
**Product Code:** 17139

Cat. No.: 1 7139 99 10 930

---

**Delivery of kits:** No of kits: 4  
Kit Lot/Los: 60134739  
Kit Expiry: 2021-03

---

**Delivery of bulk:** Volume:  
Processing Lot:

---

### Results for:

Reagent 1 Lot No.: 27044  
Expiry: 2021-03

Reagent 2 Lot No.: 27044  
Expiry: 2021-03

### Examination:

Visual check Reagent 1: clear to slightly opalescent,  
yellowish  
Reagent 2: milky, white

Smell Reagent 1: no characteristic odour  
Reagent 2: no characteristic odour

**Control sera:** TruLab Lp(a) 1 TruLab Lp(a) 2  
Lot: 26192 Lot: 26193

Test results 42.7 nmol/L 106 nmol/L

Target values 39.7 31.8 - 47.6 105 84.0 - 126

---

### Additional remarks:

This lot meets our specifications.  
This certificate of analysis is a computer printout and has therefore not been signed  
by hand.

Sign.: Dr. G. Gorka